

MAXIMUM TEMPERATURE RANGE

Thermocouple Grade

32 to 2642°F

0 to 1450°C

Extension Grade

32 to 300°F

0 to 150°C

LIMITS OF ERROR

(whichever is greater)

Standard: 1.5°C or 0.25%

Special: 0.6°C or 0.1%


COMMENTS, BARE WIRE ENVIRONMENT:

Oxidizing or Inert; Do Not Insert in Metal Tubes;

Beware of Contamination; High Temperature

TEMPERATURE IN DEGREES °F


PREFERENCE JUNCTION AT 32°F



Thermocouple Grade

NONE ESTABLISHED

Platinum-10% Rhodium vs. Platinum



Extension Grade

Revised Thermocouple Reference Tables

TYPE
Reference Tables
N.I.S.T.
Monograph 175
Revised to ITS-90



Thermoelectric Voltage in Millivolts


°F	-10	-9	-8	-7	-6	-5	-4	-3	-2	-1	0	°F
-50			-0.236	-0.233	-0.231	-0.229	-0.227	-0.224	-0.222	-0.220	-0.218	-50
-40	-0.218	-0.215	-0.213	-0.211	-0.208	-0.206	-0.204	-0.201	-0.199	-0.197	-0.194	-40
-30	-0.194	-0.192	-0.190	-0.187	-0.185	-0.182	-0.180	-0.178	-0.175	-0.173	-0.170	-30
-20	-0.170	-0.168	-0.165	-0.163	-0.160	-0.158	-0.155	-0.153	-0.150	-0.148	-0.145	-20
-10	-0.145	-0.142	-0.140	-0.137	-0.135	-0.132	-0.129	-0.127	-0.124	-0.122	-0.119	-10
0	-0.119	-0.116	-0.114	-0.111	-0.108	-0.106	-0.103	-0.100	-0.097	-0.095	-0.092	0
0	-0.092	-0.089	-0.086	-0.084	-0.081	-0.078	-0.075	-0.073	-0.070	-0.067	-0.064	0
10	-0.064	-0.061	-0.058	-0.056	-0.053	-0.050	-0.047	-0.044	-0.041	-0.038	-0.035	10
20	-0.035	-0.033	-0.030	-0.027	-0.024	-0.021	-0.018	-0.015	-0.012	-0.009	-0.006	20
30	-0.006	-0.003	0.000	0.003	0.006	0.009	0.012	0.015	0.018	0.021	0.024	30
40	0.024	0.027	0.030	0.033	0.037	0.040	0.043	0.046	0.049	0.052	0.055	40
50	0.055	0.058	0.062	0.065	0.068	0.071	0.074	0.077	0.081	0.084	0.087	50
60	0.087	0.090	0.093	0.097	0.100	0.103	0.106	0.110	0.113	0.116	0.119	60
70	0.119	0.123	0.126	0.129	0.133	0.136	0.139	0.143	0.146	0.149	0.153	70
80	0.153	0.156	0.159	0.163	0.166	0.169	0.173	0.176	0.180	0.183	0.186	80
90	0.186	0.190	0.193	0.197	0.200	0.204	0.207	0.210	0.214	0.217	0.221	90
100	0.221	0.224	0.228	0.231	0.235	0.238	0.242	0.245	0.249	0.252	0.256	100
110	0.256	0.260	0.263	0.267	0.270	0.274	0.277	0.281	0.285	0.288	0.292	110
120	0.292	0.295	0.299	0.303	0.306	0.310	0.313	0.317	0.321	0.324	0.328	120
130	0.328	0.332	0.335	0.339	0.343	0.346	0.350	0.354	0.357	0.361	0.365	130
140	0.365	0.369	0.372	0.376	0.380	0.384	0.387	0.391	0.395	0.399	0.402	140
150	0.402	0.406	0.410	0.414	0.417	0.421	0.425	0.429	0.433	0.436	0.440	150
160	0.440	0.444	0.448	0.452	0.456	0.459	0.463	0.467	0.471	0.475	0.479	160
170	0.479	0.483	0.487	0.490	0.494	0.498	0.502	0.506	0.510	0.514	0.518	170
180	0.518	0.522	0.526	0.530	0.534	0.538	0.541	0.545	0.549	0.553	0.557	180
190	0.557	0.561	0.565	0.569	0.573	0.577	0.581	0.585	0.589	0.593	0.597	190
200	0.597	0.601	0.605	0.609	0.613	0.617	0.622	0.626	0.630	0.634	0.638	200
210	0.638	0.642	0.646	0.650	0.654	0.658	0.662	0.666	0.670	0.675	0.679	210
220	0.679	0.683	0.687	0.691	0.695	0.699	0.703	0.708	0.712	0.716	0.720	220
230	0.720	0.724	0.728	0.732	0.737	0.741	0.745	0.749	0.753	0.758	0.762	230
240	0.762	0.766	0.770	0.774	0.779	0.783	0.787	0.791	0.795	0.800	0.804	240
250	0.804	0.808	0.812	0.817	0.821	0.825	0.829	0.834	0.838	0.842	0.847	250
260	0.847	0.851	0.855	0.859	0.864	0.868	0.872	0.877	0.881	0.885	0.889	260
270	0.889	0.894	0.898	0.902	0.907	0.911	0.915	0.920	0.924	0.928	0.933	270
280	0.933	0.937	0.942	0.946	0.950	0.955	0.959	0.963	0.968	0.972	0.977	280
290	0.977	0.981	0.985	0.990	0.994	0.998	1.003	1.007	1.012	1.016	1.021	290
300	1.021	1.025	1.029	1.034	1.038	1.043	1.047	1.052	1.056	1.061	1.065	300
310	1.065	1.069	1.074	1.078	1.083	1.087	1.092	1.096	1.101	1.105	1.110	310
320	1.110	1.114	1.119	1.123	1.128	1.132	1.137	1.141	1.146	1.150	1.155	320
330	1.155	1.159	1.164	1.168	1.173	1.177	1.182	1.186	1.191	1.196	1.200	330
340	1.200	1.205	1.209	1.214	1.218	1.223	1.227	1.232	1.237	1.241	1.246	340
350	1.246	1.250	1.255	1.260	1.264	1.269	1.273	1.278	1.283	1.287	1.292	350
360	1.292	1.296	1.301	1.306	1.310	1.315	1.319	1.324	1.329	1.333	1.338	360
370	1.338	1.343	1.347	1.352	1.357	1.361	1.366	1.371	1.375	1.380	1.385	370
380	1.385	1.389	1.394	1.399	1.403	1.408	1.413	1.417	1.422	1.427	1.431	380
390	1.431	1.436	1.441	1.445	1.450	1.455	1.460	1.464	1.469	1.474	1.478	390
°F	0	1	2	3	4	5	6	7	8	9	10	°F

°F	0	1	2	3	4	5	6	7	8	9	10	°F
400	1.478	1.483	1.488	1.493	1.497	1.502	1.507	1.512	1.516	1.521	1.526	400
410	1.526	1.531	1.535	1.540	1.545	1.550	1.554	1.559	1.564	1.569	1.573	410
420	1.573	1.578	1.583	1.588	1.592	1.597	1.602	1.607	1.612	1.616	1.621	420
430	1.621	1.626	1.631	1.636	1.640	1.645	1.650	1.655	1.660	1.664	1.669	430
440	1.669	1.674	1.679	1.684	1.689	1.693	1.698	1.703	1.708	1.713	1.718	440
450	1.718	1.722	1.727	1.732	1.737	1.742	1.747	1.752	1.756	1.761	1.766	450
460	1.766	1.771	1.776	1.781	1.786	1.790	1.795	1.800	1.805	1.810	1.815	460
470	1.815	1.820	1.825	1.829	1.834	1.839	1.844	1.849	1.854	1.859	1.864	470
480	1.864	1.869	1.874	1.878	1.883	1.888	1.893	1.898	1.903	1.908	1.913	480
490	1.913	1.918	1.923	1.928	1.933	1.938	1.942	1.947	1.952	1.957	1.962	490
500	1.962	1.967	1.972	1.977	1.982	1.987	1.992	1.997	2.002	2.007	2.012	500
510	2.012	2.017	2.022	2.027	2.032	2.037	2.042	2.047	2.052	2.057	2.062	510
520	2.062	2.067	2.072	2.076	2.081	2.086	2.091	2.096	2.101	2.106	2.111	520
530	2.111	2.116	2.121	2.126	2.131	2.136	2.141	2.147	2.152	2.157	2.162	530
540	2.162	2.167	2.172	2.177	2.182	2.187	2.192	2.197	2.202	2.207	2.212	540
550	2.212	2.217	2.222	2.227	2.232	2.237	2.242	2.247	2.252	2.257	2.262	550
560	2.262	2.267	2.272	2.277	2.283	2.288	2.293	2.298	2.303	2.308	2.313	560
570	2.313	2.318	2.323	2.328	2.333	2.338	2.343	2.348	2.354	2.359	2.364	570
580	2.364	2.369	2.374	2.379	2.384	2.389	2.394	2.399	2.404	2.409	2.415	580
590	2.415	2.420	2.425	2.430	2.435	2.440	2.445	2.450	2.455	2.461	2.466	590
600	2.466	2.471	2.476	2.481	2.486	2.491	2.496	2.502	2.507	2.512	2.517	600
610	2.517	2.522	2.527	2.532	2.537	2.543	2.548	2.553	2.558	2.563	2.568	610
620	2.568	2.574	2.579	2.584	2.589	2.594	2.599	2.604	2.610	2.615	2.620	620
630	2.620	2.625	2.630	2.635	2.641	2.646	2.651	2.656	2.661	2.666	2.672	630
640	2.672	2.677	2.682	2.687	2.692	2.697	2.703	2.708	2.713	2.718	2.723	640
650	2.723	2.729	2.734	2.739	2.744	2.749	2.755	2.760	2.765	2.770	2.775	650
660	2.775	2.781	2.786	2.791	2.796	2.801	2.807	2.812	2.817	2.822	2.827	660
670	2.827	2.833	2.838	2.843	2.848	2.854	2.859	2.864	2.869	2.874	2.880	670
680	2.880	2.885	2.890	2.895	2.901	2.906	2.911	2.916	2.922	2.927	2.932	680
690	2.932	2.937	2.943	2.948	2.953	2.958	2.964	2.969	2.974	2.979	2.985	690
700	2.985	2.990	2.995	3.000	3.006	3.011	3.016	3.021	3.027	3.032	3.037	700
710	3.037	3.042	3.048	3.053	3.058	3.063	3.069	3.074	3.079	3.085	3.090	710
720	3.090	3.095	3.100	3.106	3.111	3.116	3.122	3.127	3.132	3.137	3.143	720
730	3.143	3.148	3.153	3.159	3.164	3.169	3.174	3.180	3.185	3.190	3.196	730
740	3.196	3.201	3.206	3.212	3.217	3.222	3.227	3.233	3.238	3.243	3.249	740
750	3.249	3.254	3.259	3.265	3.270	3.275	3.281	3.286	3.291	3.297	3.302	750
760	3.302	3.307	3.313	3.318	3.323	3.329	3.334	3.339	3.345	3.350	3.355	760
770	3.355	3.361	3.366</									

Revised Thermocouple Reference Tables

TYPE S


Reference Tables
N.I.S.T.
Monograph 175
Revised to
ITS-90



Thermocouple Grade
NONE ESTABLISHED

Platinum-10% Rhodium
VS.
Platinum

Extension Grade



MAXIMUM TEMPERATURE RANGE

Thermocouple Grade

32 to 2642°F
0 to 1450°C

Extension Grade

32 to 300°F
0 to 150°C

LIMITS OF ERROR

(whichever is greater)

Standard: 1.5°C or 0.25%

Special: 0.6°C or 0.1%

COMMENTS, BARE WIRE ENVIRONMENT:
Oxidizing or Inert; Do Not Insert in Metal Tubes;
Beware of Contamination; High Temperature

TEMPERATURE IN DEGREES °F
PREFERENCE JUNCTION AT 32°F

Thermoelectric Voltage in Millivolts

°F	-10	-9	-8	-7	-6	-5	-4	-3	-2	-1	0	°F
900	4.058	4.063	4.069	4.074	4.080	4.085	4.091	4.096	4.102	4.107	4.113	900
910	4.113	4.118	4.123	4.129	4.134	4.140	4.145	4.151	4.156	4.162	4.167	910
920	4.167	4.173	4.178	4.184	4.189	4.195	4.200	4.206	4.211	4.217	4.222	920
930	4.222	4.228	4.233	4.239	4.244	4.250	4.255	4.261	4.266	4.272	4.277	930
940	4.277	4.283	4.288	4.294	4.299	4.305	4.310	4.316	4.321	4.327	4.332	940
950	4.332	4.338	4.343	4.349	4.355	4.360	4.366	4.371	4.377	4.382	4.388	950
960	4.388	4.393	4.399	4.404	4.410	4.415	4.421	4.426	4.432	4.437	4.443	960
970	4.443	4.449	4.454	4.460	4.465	4.471	4.476	4.482	4.487	4.493	4.498	970
980	4.498	4.504	4.510	4.515	4.521	4.526	4.532	4.537	4.543	4.548	4.554	980
990	4.554	4.559	4.565	4.571	4.576	4.582	4.587	4.593	4.598	4.604	4.610	990
1000	4.610	4.615	4.621	4.626	4.632	4.637	4.643	4.648	4.654	4.660	4.665	1000
1010	4.665	4.671	4.676	4.682	4.688	4.693	4.699	4.704	4.710	4.715	4.721	1010
1020	4.721	4.727	4.732	4.738	4.743	4.749	4.755	4.760	4.766	4.771	4.777	1020
1030	4.777	4.782	4.788	4.794	4.799	4.805	4.810	4.816	4.822	4.827	4.833	1030
1040	4.833	4.838	4.844	4.850	4.855	4.861	4.866	4.872	4.878	4.883	4.889	1040
1050	4.889	4.895	4.900	4.906	4.911	4.917	4.923	4.928	4.934	4.939	4.945	1050
1060	4.945	4.951	4.956	4.962	4.968	4.973	4.979	4.984	4.990	4.996	5.001	1060
1070	5.001	5.007	5.013	5.018	5.024	5.030	5.035	5.041	5.046	5.052	5.058	1070
1080	5.058	5.063	5.069	5.075	5.080	5.086	5.092	5.097	5.103	5.109	5.114	1080
1090	5.114	5.120	5.125	5.131	5.137	5.142	5.148	5.154	5.159	5.165	5.171	1090
1100	5.171	5.176	5.182	5.188	5.193	5.199	5.205	5.210	5.216	5.222	5.227	1100
1110	5.227	5.233	5.239	5.244	5.250	5.256	5.261	5.267	5.273	5.278	5.284	1110
1120	5.284	5.290	5.295	5.301	5.307	5.312	5.318	5.324	5.330	5.335	5.341	1120
1130	5.341	5.347	5.352	5.358	5.364	5.369	5.375	5.381	5.386	5.392	5.398	1130
1140	5.398	5.404	5.409	5.415	5.421	5.426	5.432	5.438	5.443	5.449	5.455	1140
1150	5.455	5.461	5.466	5.472	5.478	5.483	5.489	5.495	5.501	5.506	5.512	1150
1160	5.512	5.518	5.523	5.529	5.535	5.541	5.546	5.552	5.558	5.563	5.569	1160
1170	5.569	5.575	5.581	5.586	5.592	5.598	5.604	5.609	5.615	5.621	5.627	1170
1180	5.627	5.632	5.638	5.644	5.649	5.655	5.661	5.667	5.672	5.678	5.684	1180
1190	5.684	5.690	5.695	5.701	5.707	5.713	5.718	5.724	5.730	5.736	5.741	1190
1200	5.741	5.747	5.753	5.759	5.764	5.770	5.776	5.782	5.788	5.793	5.799	1200
1210	5.799	5.805	5.811	5.816	5.822	5.828	5.834	5.839	5.845	5.851	5.857	1210
1220	5.857	5.863	5.868	5.874	5.880	5.886	5.891	5.897	5.903	5.909	5.915	1220
1230	5.915	5.920	5.926	5.932	5.938	5.944	5.949	5.955	5.961	5.967	5.972	1230
1240	5.972	5.978	5.984	5.990	5.996	6.001	6.007	6.013	6.019	6.025	6.030	1240
1250	6.030	6.036	6.042	6.048	6.054	6.060	6.065	6.071	6.077	6.083	6.089	1250
1260	6.089	6.094	6.100	6.106	6.112	6.118	6.124	6.129	6.135	6.141	6.147	1260
1270	6.147	6.153	6.158	6.164	6.170	6.176	6.182	6.188	6.193	6.199	6.205	1270
1280	6.205	6.211	6.217	6.223	6.228	6.234	6.240	6.246	6.252	6.258	6.264	1280
1290	6.264	6.269	6.275	6.281	6.287	6.293	6.299	6.305	6.310	6.316	6.322	1290
1300	6.322	6.328	6.334	6.340	6.346	6.351	6.357	6.363	6.369	6.375	6.381	1300
1310	6.381	6.387	6.393	6.398	6.404	6.410	6.416	6.422	6.428	6.434	6.439	1310
1320	6.439	6.445	6.451	6.457	6.463	6.469	6.475	6.481	6.486	6.492	6.498	1320
1330	6.498	6.504	6.510	6.516	6.522	6.528	6.534	6.539	6.545	6.551	6.557	1330
1340	6.557	6.563	6.569	6.575	6.581	6.587	6.593	6.598	6.604	6.610	6.616	1340
1350	6.616	6.622	6.628	6.634	6.640	6.646	6.652	6.658	6.664	6.669	6.675	1350
1360	6.675	6.681	6.687	6.693	6.699	6.705	6.711	6.717	6.723	6.729	6.735	1360
1370	6.735	6.741	6.746	6.752	6.758	6.764	6.770	6.776	6.782	6.788	6.794	1370
1380	6.794	6.800	6.806	6.812	6.818	6.824	6.830	6.836	6.842	6.847	6.853	1380
1390	6.853	6.859	6.865	6.871	6.877	6.883	6.889	6.895	6.901	6.907	6.913	1390

°F	0	1	2	3	4	5	6	7	8	9	10	°F
1400	6.913	6.919	6.925	6.931	6.937	6.943	6.949	6.955	6.961	6.967	6.973	1400
1410	6.973	6.979	6.985	6.991	6.997	7.003	7.008	7.014	7.020	7.026	7.032	1410
1420	7.032	7.038	7.044	7.050	7.056	7.062	7.068	7.074	7.080	7.086	7.092	1420
1430	7.092	7.098	7.104	7.110	7.116	7.122	7.128	7.134	7.140	7.146	7.152	1430
1440	7.152	7.158	7.164	7.170	7.176	7.182	7.188	7.194	7.200	7.206	7.212	1440
1450	7.212	7.218	7.224	7.230	7.236	7.242	7.249	7.255	7.261	7.267	7.273	1450
1460	7.273	7.279	7.285	7.291	7.297	7.303	7.309	7.315	7.321	7.327	7.333	1460
1470	7.333	7.339	7.345	7.351	7.357	7.363	7.369	7.375	7.381	7.387	7.393	1470
1480	7.393	7.399	7.405	7.411	7.417	7.424	7.430	7.436	7.442	7.448	7.454	1480
1490	7.454	7.460	7.466	7.472	7.478	7.484	7.490	7.496	7.502	7.508	7.514	1490
1500	7.514	7.521	7.527	7.533	7.539	7.545	7.551	7.557	7.563	7.569	7.575	1500
1510	7.575	7.581	7.587	7.593	7.600	7.606	7.612	7.618	7.624	7.630	7.636	1510
1520	7.636	7.642	7.648	7.654	7.660	7.667	7.673	7.679	7.685	7.691	7.697	1520
1530	7.697	7.703	7.709	7.715	7.721	7.728	7.734	7.740	7.746	7.752	7.758	1530
1540	7.758	7.764	7.770	7.776	7.782	7.789	7.795	7.801	7.807	7.813	7.819	1540
1550	7.819	7.825	7.832	7.838	7.844	7.850	7.856	7.862	7.868	7.874	7.881	1550
1560	7.881	7.887	7.893	7.899	7.905	7.911	7.917	7.923	7.930	7.936	7.942	1560
1570	7.942	7.948	7.954	7.960	7.966	7.973	7.979	7.985	7.991	7.997	8.003	1570
1580	8.003	8.010	8.016	8.022	8.028	8.034	8.040	8.047	8.053	8.059	8.065	1580
1590	8.065	8.071	8.077	8.083	8.090	8.096	8.102	8.108	8.114	8.121	8.127	1590
1600	8.127	8.133	8.139	8.145	8.151	8.158	8.164	8.170	8.176	8.182	8.189	1600
1610	8.189	8.195	8.201	8.207	8.213	8.219	8.226	8.232	8.238	8.244	8.250	1610
1620	8.250	8.257	8.263	8.269	8.275	8.281	8.288	8.294	8.300	8.306	8.312	1620
1630	8.312	8.319	8.325	8.331	8.337	8.343	8.350	8.356	8.362	8.368	8.375	1630
1640	8.375	8.381	8.387	8.393	8.399	8.406	8.412	8.418	8.424	8.431	8.437	1640
1650	8.437	8.443	8.449	8.455	8.462	8.468	8.474	8.480	8.487	8.493	8.499	1650
1660	8.499	8.505	8.512	8.518	8.524	8.530	8.537	8.543	8.549	8.555	8.562	1660
1670	8.562	8.568	8.574	8.580	8.587	8.593	8.599	8.605	8.612	8.618	8.624	1670
1680	8.624	8.630	8.637	8.643	8.649	8.655	8.662	8.668	8.674	8.680	8.687	1680
1690	8.687	8.693	8.699	8.706	8.712	8.718	8.724	8.731	8.737	8.743	8.749	1690
1700	8.749	8.756	8.762	8.768	8.775	8.781	8.787	8.793	8.800	8.806	8.812	1700
1710	8.812	8.819	8.825	8.831	8.837	8.844	8.850	8.856	8.863	8.869	8.875	1710
1720	8.875	8.882	8.888	8.894	8.900	8.907	8.913	8.919	8.926	8.932	8.938	1720
1730	8.938	8.945	8.									

MAXIMUM TEMPERATURE RANGE

Thermocouple Grade

32 to 2642°F
0 to 1450°C

Extension Grade


32 to 300°F
0 to 150°C

LIMITS OF ERROR
(whichever is greater)

Standard: 1.5°C or 0.25%
Special: 0.6°C or 0.1%

COMMENTS, BARE WIRE ENVIRONMENT:
Oxidizing or Inert; Do Not Insert in Metal Tubes;
Beware of Contamination; High Temperature


TEMPERATURE IN DEGREES °F
REFERENCE JUNCTION AT 32°F



Thermocouple Grade

NONE ESTABLISHED

Platinum-10% Rhodium
vs.
Platinum



Extension Grade

Revised Thermocouple Reference Tables

TYPE S

Reference Tables
N.I.S.T.
Monograph 175
Revised to ITS-90


Thermoelectric Voltage in Millivolts

°F	0	1	2	3	4	5	6	7	8	9	10	°F	0	1	2	3	4	5	6	7	8	9	10	°F				
1900	10.025	10.032	10.038	10.045	10.051	10.058	10.064	10.071	10.077	10.084	10.090	1900	13.400	13.408	13.415	2400	13.348	13.354	13.361	13.368	13.375	13.381	13.388	13.395	13.402	13.408	13.415	2400
1910	10.090	10.097	10.103	10.110	10.116	10.123	10.129	10.136	10.142	10.149	10.155	1910	13.415	13.422	13.429	2410	13.415	13.422	13.429	13.435	13.442	13.449	13.456	13.462	13.469	13.476	13.483	2410
1920	10.155	10.162	10.168	10.175	10.181	10.188	10.194	10.201	10.207	10.214	10.220	1920	13.420	13.427	13.434	2420	13.483	13.489	13.496	13.503	13.510	13.516	13.523	13.530	13.537	13.543	13.550	2420
1930	10.220	10.227	10.233	10.240	10.246	10.253	10.259	10.266	10.272	10.279	10.285	1930	13.550	13.557	13.563	2430	13.550	13.557	13.563	13.570	13.577	13.584	13.590	13.597	13.604	13.611	13.617	2430
1940	10.285	10.292	10.298	10.305	10.311	10.318	10.324	10.331	10.337	10.344	10.350	1940	13.617	13.624	13.631	2440	13.617	13.624	13.631	13.638	13.644	13.651	13.658	13.665	13.671	13.678	13.685	2440
1950	10.350	10.357	10.363	10.370	10.376	10.383	10.390	10.396	10.403	10.409	10.416	1950	13.685	13.692	13.698	2450	13.685	13.692	13.698	13.705	13.712	13.719	13.725	13.732	13.739	13.746	13.752	2450
1960	10.416	10.422	10.429	10.435	10.442	10.448	10.455	10.461	10.468	10.475	10.481	1960	13.752	13.759	13.766	2460	13.752	13.759	13.766	13.773	13.779	13.786	13.793	13.800	13.806	13.813	13.820	2460
1970	10.481	10.488	10.494	10.501	10.507	10.514	10.520	10.527	10.533	10.540	10.547	1970	13.820	13.826	13.833	2470	13.820	13.826	13.833	13.840	13.847	13.853	13.860	13.867	13.874	13.880	13.887	2470
1980	10.547	10.553	10.560	10.566	10.573	10.579	10.586	10.592	10.599	10.606	10.612	1980	13.887	13.893	13.901	2480	13.887	13.893	13.901	13.907	13.914	13.921	13.928	13.934	13.941	13.948	13.955	2480
1990	10.612	10.619	10.625	10.632	10.638	10.645	10.651	10.658	10.665	10.671	10.678	1990	13.955	13.961	13.968	2490	13.955	13.961	13.968	13.975	13.982	13.988	13.995	14.002	14.009	14.015	14.022	2490
2000	10.678	10.684	10.691	10.697	10.704	10.711	10.717	10.724	10.730	10.737	10.743	2000	14.022	14.029	14.036	2500	14.022	14.029	14.036	14.042	14.049	14.056	14.063	14.069	14.076	14.083	14.089	2500
2010	10.743	10.750	10.757	10.763	10.770	10.776	10.783	10.789	10.796	10.803	10.809	2010	14.089	14.096	14.103	2510	14.089	14.096	14.103	14.110	14.116	14.123	14.130	14.137	14.143	14.150	14.157	2510
2020	10.809	10.816	10.822	10.829	10.836	10.842	10.849	10.855	10.862	10.868	10.875	2020	14.157	14.164	14.170	2520	14.157	14.164	14.170	14.177	14.184	14.191	14.197	14.204	14.211	14.218	14.224	2520
2030	10.875	10.882	10.888	10.895	10.901	10.908	10.915	10.921	10.928	10.934	10.941	2030	14.224	14.231	14.238	2530	14.224	14.231	14.238	14.245	14.251	14.258	14.265	14.272	14.278	14.285	14.292	2530
2040	10.941	10.948	10.954	10.961	10.967	10.974	10.981	10.987	10.994	11.000	11.007	2040	14.292	14.298	14.305	2540	14.292	14.298	14.305	14.312	14.319	14.325	14.332	14.339	14.346	14.352	14.359	2540
2050	11.007	11.014	11.020	11.027	11.033	11.040	11.047	11.053	11.060	11.066	11.073	2050	14.359	14.366	14.373	2550	14.359	14.366	14.373	14.379	14.386	14.393	14.400	14.406	14.413	14.420	14.426	2550
2060	11.073	11.080	11.086	11.093	11.099	11.106	11.113	11.119	11.126	11.132	11.139	2060	14.426	14.433	14.440	2560	14.426	14.433	14.440	14.447	14.453	14.460	14.467	14.474	14.480	14.487	14.494	2560
2070	11.139	11.146	11.152	11.159	11.166	11.172	11.179	11.185	11.192	11.199	11.205	2070	14.494	14.501	14.507	2570	14.494	14.501	14.507	14.514	14.521	14.528	14.534	14.541	14.548	14.554	14.561	2570
2080	11.205	11.212	11.219	11.225	11.232	11.238	11.245	11.252	11.258	11.265	11.272	2080	14.561	14.568	14.575	2580	14.561	14.568	14.575	14.581	14.588	14.595	14.602	14.608	14.615	14.622	14.629	2580
2090	11.272	11.278	11.285	11.291	11.298	11.305	11.311	11.318	11.325	11.331	11.338	2090	14.629	14.635	14.642	2590	14.629	14.635	14.642	14.649	14.655	14.662	14.669	14.676	14.682	14.689	14.696	2590
2100	11.338	11.345	11.351	11.358	11.364	11.371	11.378	11.384	11.391	11.398	11.404	2100	14.696	14.703	14.709	2600	14.696	14.703	14.709	14.716	14.723	14.729	14.736	14.743	14.750	14.756	14.763	2600
2110	11.404	11.411	11.418	11.424	11.431	11.437	11.444	11.451	11.457	11.464	11.471	2110	14.763	14.770	14.777	2610	14.763	14.770	14.777	14.783	14.790	14.797	14.803	14.810	14.817	14.824	14.830	2610
2120	11.471	11.477	11.484	11.491	11.497	11.504	11.511	11.517	11.524	11.531	11.537	2120	14.830	14.837	14.844	2620	14.830	14.837	14.844	14.851	14.857	14.864	14.871	14.877	14.884	14.891	14.898	2620
2130	11.537	11.544	11.550	11.557	11.564	11.570	11.577	11.584	11.590	11.597	11.604	2130	14.898	14.904	14.911	2630	14.898	14.904	14.911	14.918	14.925	14.931	14.938	14.945	14.951	14.958	14.965	2630
2140	11.604	11.610	11.617	11.624	11.631	11.637	11.644	11.650	11.657	11.664	11.670	2140	14.965	14.972	14.978	2640	14.965	14.972	14.978	14.985	14.992	14.998	15.005	15.012	15.019	15.025	15.032	2640
2150	11.670	11.677	11.684	11.690	11.697	11.704	11.710	11.717	11.724	11.730	11.737	2150	15.032	15.039	15.045	2650	15.032	15.039	15.045	15.052	15.059	15.066	15.072	15.079	15.086	15.092	15.099	2650
2160	11.737	11.744	11.750	11.757	11.764	11.770	11.777	11.784	11.790	11.797	11.804	2160	15.099	15.106	15.113	2660	15.099	15.106	15.113	15.119	15.126	15.133	15.139	15.146	15.153	15.160	15.166	2660
2170	11.804	11.810	11.817	11.824	11.830	11.837	11.844	11.850	11.857	11.864	11.870	2170	15.166	15.173	15.180	2670	15.166	15.173	15.180	15.186	15.193	15.200	15.207	15.213	15.220	15.227	15.233	2670
2180	11.870	11.877	11.884	11.890	11.897	11.904	11.910	11.917	11.924	11.931	11.937	2180	15.233	15.240	15.247	2680	15.233	15.240	15.247	15.254	15.260	15.267	15.274	15.280	15.287	15.294	15.300	2680
2190	11.937	11.944	11.951	11.957	11.964	11.971	11.977	11.984	11.991	11.997	12.004	2190	15.300	15.307	15.314	2690	15.300	15.307	15.314	15.321	15.327	15.334	15.341	15.347	15.354	15.361	15.367	2690
2200	12.004	12.011	12.017	12.024	12.031	12.037	12.044	12.051	12.058	12.064	12.071	2200	15.367	15.374	15.381	2700	15.367	15.374	15.381	15.388	15.394	15.401	15.408	15.414	15.421	15.428	15.434	2700
2210	12.071	12.078	12.084	12.091	12.098	12.104	12.111	12.118	12.124	12.131	12.138	2210	15.434	15.441	15.448	2710	15.434	15.441	15.448	15.455	15.461	15.468	15.475	15.481	15.488	15.495	15.501	2710
2220	12.138	12.145	12.151	12.158	12.165	12.171	12.178	12.185	12.191	12.198	12.205	2220	15.501	15.508	15.515	2720	15.501	15.508	15.515	15.521	15.528	15.535	15.542	15.548	15.555	15.562	15.568	2720
2230	12.205	12.211	12.218	12.225	12.232	12.238	12.245	12.252	12.258	12.265	12.272	2230	15.568	15.575	15.582	2730	15.568	15.575	15.582	15.588	15.595	15.602	15.608	15.615	15.622	15.628	15.635	2730
2240	12.272	12.278	12.285	12.292	12.299	12.305	12.312	12.319	12.325	12.332	12.339	2240	15.635	15.642	15.649	2740	15.635	15.642	15.649	15.655	15.662	15.669	15.675	15.682	15.689	15.695	15.702	2740
2250	12.339	12.346	12.352	12.359	12.366	12.372	12.379	12.386	12.392	12.399	12.406	2250	15.702	15.709	15.715	2750	15.702											

Revised Thermocouple Reference Tables

TYPE S


Reference Tables
N.I.S.T.
Monograph 175
Revised to ITS-90



Thermocouple Grade

NONE ESTABLISHED

Platinum-10% Rhodium
VS.
Platinum



Extension Grade

MAXIMUM TEMPERATURE RANGE

Thermocouple Grade

32 to 2642°F
0 to 1450°C

Extension Grade

32 to 300°F
0 to 150°C

LIMITS OF ERROR
(whichever is greater)

Standard: 1.5°C or 0.25%

Special: 0.6°C or 0.1%

COMMENTS, BARE WIRE ENVIRONMENT:
Oxidizing or Inert; Do Not Insert in Metal Tubes;
Beware of Contamination; High Temperature

TEMPERATURE IN DEGREES °F
REFERENCE JUNCTION AT 32°F

Thermoelectric Voltage in Millivolts

°F	0	1	2	3	4	5	6	7	8	9	10	°F	°F	0	1	2	3	4	5	6	7	8	9	10	°F
2900	16.698	16.704	16.711	16.718	16.724	16.731	16.737	16.744	16.751	16.757	16.764	2900	3100	17.998	18.004	18.011	18.017	18.023	18.030	18.036	18.042	18.049	18.055	18.061	3100
2910	16.764	16.770	16.777	16.783	16.790	16.797	16.803	16.810	16.816	16.823	16.829	2910	3110	18.061	18.068	18.074	18.080	18.086	18.093	18.099	18.105	18.112	18.118	18.124	3110
2920	16.829	16.836	16.843	16.849	16.856	16.862	16.869	16.876	16.882	16.889	16.895	2920	3120	18.124	18.130	18.137	18.143	18.149	18.155	18.162	18.168	18.174	18.180	18.187	3120
2930	16.895	16.902	16.908	16.915	16.922	16.928	16.935	16.941	16.948	16.954	16.961	2930	3130	18.187	18.193	18.199	18.205	18.211	18.218	18.224	18.230	18.236	18.242	18.248	3130
2940	16.961	16.967	16.974	16.981	16.987	16.994	17.000	17.007	17.013	17.020	17.026	2940	3140	18.248	18.255	18.261	18.267	18.273	18.279	18.285	18.292	18.298	18.304	18.310	3140
2950	17.026	17.033	17.040	17.046	17.053	17.059	17.066	17.072	17.079	17.085	17.092	2950	3150	18.310	18.316	18.322	18.328	18.334	18.341	18.347	18.353	18.359	18.365	18.371	3150
2960	17.092	17.099	17.105	17.112	17.118	17.125	17.131	17.138	17.144	17.151	17.157	2960	3160	18.371	18.377	18.383	18.389	18.395	18.401	18.407	18.413	18.419	18.425	18.431	3160
2970	17.157	17.164	17.171	17.177	17.184	17.190	17.197	17.203	17.210	17.216	17.223	2970	3170	18.431	18.437	18.443	18.449	18.455	18.461	18.467	18.473	18.479	18.485	18.491	3170
2980	17.223	17.229	17.236	17.242	17.249	17.255	17.262	17.268	17.275	17.282	17.288	2980	3180	18.491	18.497	18.503	18.509	18.515	18.521	18.527	18.533	18.539	18.545	18.551	3180
2990	17.288	17.295	17.301	17.308	17.314	17.321	17.327	17.334	17.340	17.347	17.353	2990	3190	18.551	18.557	18.562	18.568	18.574	18.580	18.586	18.592	18.598	18.603	18.609	3190
3000	17.353	17.360	17.366	17.373	17.379	17.386	17.392	17.399	17.405	17.412	17.418	3000	3200	18.609	18.615	18.621	18.627	18.633	18.638	18.644	18.650	18.656	18.661	18.667	3200
3010	17.418	17.425	17.431	17.438	17.444	17.451	17.457	17.464	17.470	17.477	17.483	3010	3210	18.667	18.673	18.679	18.684	18.690							3210
3020	17.483	17.490	17.496	17.503	17.509	17.516	17.522	17.529	17.535	17.542	17.548	3020													
3030	17.548	17.555	17.561	17.568	17.574	17.581	17.587	17.594	17.600	17.607	17.613	3030													
3040	17.613	17.620	17.626	17.633	17.639	17.645	17.652	17.658	17.665	17.671	17.678	3040													
3050	17.678	17.684	17.691	17.697	17.704	17.710	17.717	17.723	17.729	17.736	17.742	3050													
3060	17.742	17.749	17.755	17.762	17.768	17.775	17.781	17.787	17.794	17.800	17.807	3060													
3070	17.807	17.813	17.819	17.826	17.832	17.839	17.845	17.852	17.858	17.864	17.871	3070													
3080	17.871	17.877	17.884	17.890	17.896	17.903	17.909	17.915	17.922	17.928	17.935	3080													
3090	17.935	17.941	17.947	17.954	17.960	17.966	17.973	17.979	17.985	17.992	17.998	3090													
°F	0	1	2	3	4	5	6	7	8	9	10	°F	°F	0	1	2	3	4	5	6	7	8	9	10	°F



UNITED STATES

www.omega.com
1-800-TC-OMEGA
Stamford, CT.

CANADA

www.omega.ca
Laval(Quebec)
1-800-TC-OMEGA

GERMANY

www.omega.de
Deckenpfronn, Germany
0800-8266342

UNITED KINGDOM

www.omega.co.uk
Manchester, England
0800-488-488

FRANCE

www.omega.fr
Guyancourt, France
088-466-342

CZECH REPUBLIC

www.omegaeng.cz
Karviná, Czech Republic
596-311-899

BENELUX

www.omega.nl
Amstelveen, NL
0800-099-33-44



More than 100,000 Products Available!

• Temperature

Calibrators, Connectors, General Test and Measurement Instruments, Glass Bulb Thermometers, Handheld Instruments for Temperature Measurement, Ice Point References, Indicating Labels, Crayons, Cements and Lacquers, Infrared Temperature Measurement Instruments, Recorders Relative Humidity Measurement Instruments, RTD Probes, Elements and Assemblies, Temperature & Process Meters, Timers and Counters, Temperature and Process Controllers and Power Switching Devices, Thermistor Elements, Probes and Assemblies, Thermocouples Thermowells and Head and Well Assemblies, Transmitters, Wire

• Flow and Level

Air Velocity Indicators, Doppler Flowmeters, Level Measurement, Magnetic Flowmeters, Mass Flowmeters, Pitot Tubes, Pumps, Rotameters, Turbine and Paddle Wheel Flowmeters, Ultrasonic Flowmeters, Valves, Variable Area Flowmeters, Vortex Shedding Flowmeters

• pH and Conductivity

Conductivity Instrumentation, Dissolved Oxygen Instrumentation, Environmental Instrumentation, pH Electrodes and Instruments, Water and Soil Analysis Instrumentation

• Data Acquisition

Auto-Dialers and Alarm Monitoring Systems, Communication Products and Converters, Data Acquisition and Analysis Software, Data Loggers Plug-in Cards, Signal Conditioners, USB, RS232, RS485 and Parallel Port Data Acquisition Systems, Wireless Transmitters and Receivers

• Pressure, Strain and Force

Displacement Transducers, Dynamic Measurement Force Sensors, Instrumentation for Pressure and Strain Measurements, Load Cells, Pressure Gauges, Pressure Reference Section, Pressure Switches, Pressure Transducers, Proximity Transducers, Regulators, Strain Gages, Torque Transducers, Valves

• Heaters

Band Heaters, Cartridge Heaters, Circulation Heaters, Comfort Heaters, Controllers, Meters and Switching Devices, Flexible Heaters, General Test and Measurement Instruments, Heater Hook-up Wire, Heating Cable Systems, Immersion Heaters, Process Air and Duct, Heaters, Radiant Heaters, Strip Heaters, Tubular Heaters